

Excluded In Calib		Excluded In Results	Peak Run	Excluded	Multi Element	Semi Quant	Internal Standard	Standard Addition				
Dilution Corrected Concentration Per Analyte			S-Calibration Has Edited Standard F-Interference Correction Failed		E-Calibration Edited T-Tripped	I-Invalid Calibration P-Pulse Counting		V-Valley Integration Failed M-Result Over Max				
Run	Label	TimeStamp	45Sc	47Ti	51V	52Cr	53Cr	54Fe	55Mn	59Co	60Ni	65Cu
			ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb
	SD of ICB;ICB		(P)n/a	(P)0.0422	(P)0.0131	(P)0.0078	(PI)3.0490	(P)0.5927	(P)0.0046	(P)0.0023	(P)0.0059	(P)0.0028
	%RSD of ICB;ICB		(P)n/a	(P)395.5000	(P)130.6000	(P)594.3000	(PI)76.0500	(P)600.4000	(P)72.7200	(P)73.0500	(P)670.0000	(P)10.1400
1	CRII;CRI	10/28/2005 11:43:59	(P)1.002	(P)0.0580	(P)1.0370	(P)1.9760	(PI)20.8800	(P)-0.2470	(P)0.9691	(P)0.9738	(P)0.9524	(P)2.0700
2	CRII;CRI	10/28/2005 11:44:55	(P)1.016	(P)-0.0035	(P)0.9928	(P)1.9320	(PI)10.6700	(P)-0.5696	(P)0.9797	(P)0.9714	(P)1.0490	(P)2.0060
3	CRII;CRI	10/28/2005 11:45:51	(P)1.025	(P)0.0296	(P)0.9839	(P)2.0000	(PI)7.0280	(P)-0.0660	(P)0.9858	(P)0.9720	(P)1.0170	(P)2.0050
	Mean of CRII;CRI	10/28/2005 11:43:59	(P)1.002	(P)0.0280	(P)1.0050	(P)1.9690	(PI)12.8600	(P)-0.2942	(P)0.9782	(P)0.9724	(P)1.0060	(P)2.0270
	SD of CRII;CRI		(P)n/a	(P)0.0308	(P)0.0285	(P)0.0343	(PI)7.1790	(P)0.2551	(P)0.0085	(P)0.0012	(P)0.0493	(P)0.0374
	%RSD of CRII;CRI		(P)n/a	(P)109.8000	(P)2.8320	(P)1.7420	(PI)55.8300	(P)86.7300	(P)0.8655	(P)0.1282	(P)4.8960	(P)1.8420
1	ICSAI;ICSA/5	10/28/2005 11:49:43	(P)1.114	(PM)2024.000	(P)-0.9465	(P)19.2500	(PI)-6209.000	(M)86110.000	(P)8.0300	(P)1.2850	(P)6.8240	(P)9.3100
2	ICSAI;ICSA/5	10/28/2005 11:50:39	(P)1.098	(PM)1984.000	(P)0.0225	(P)19.0300	(PI)-5855.000	(M)85900.000	(P)8.2410	(P)1.2760	(P)7.3330	(P)9.2380
3	ICSAI;ICSA/5	10/28/2005 11:51:35	(P)1.092	(PM)2010.000	(P)1.0610	(P)19.0100	(PI)-5783.000	(M)86910.000	(P)8.2500	(P)1.2370	(P)7.0620	(P)9.4460
	Mean of ICSAI;ICSA/5	10/28/2005 11:49:43	(P)1.114	(PM)2006.000	(P)0.0456	(P)19.1000	(PI)-5949.000	(M)86300.000	(P)8.1730	(P)1.2660	(P)7.0730	(P)9.3310
	SD of ICSAI;ICSA/5		(P)n/a	(PM)20.3400	(P)1.0040	(P)0.1358	(PI)227.8000	(M)532.1000	(P)0.1246	(P)0.0254	(P)0.2549	(P)0.1056
	%RSD of ICSAI;ICSA		(P)n/a	(PM)1.0140	(P)2201.0000	(P)0.7111	(PI)3.8290	(M)0.6166	(P)1.5250	(P)2.0050	(P)3.6040	(P)1.1320
1	ICSABI;ICSAB/5	10/28/2005 11:55:27	(P)1.097	(PM)2073.000	(P)15.2000	(P)38.9000	(PI)-6883.000	(M)89330.000	(P)27.5200	(P)20.5800	(P)25.4900	(P)29.6400
2	ICSABI;ICSAB/5	10/28/2005 11:56:23	(P)1.108	(PM)2076.000	(P)19.5800	(P)38.4900	(PI)-6282.000	(M)90280.000	(P)27.6200	(P)20.6200	(P)25.4500	(P)28.1700
3	ICSABI;ICSAB/5	10/28/2005 11:57:19	(P)1.090	(PM)2070.000	(P)21.5700	(P)38.8800	(PI)-6429.000	(M)90570.000	(P)27.6700	(P)20.4800	(P)24.9300	(P)27.8000
	Mean of ICSABI;ICSAB/5	10/28/2005 11:55:27	(P)1.097	(PM)2073.000	(P)18.7900	(P)38.7600	(PI)-6531.000	(M)90060.000	(P)27.6000	(P)20.5600	(P)25.2900	(P)28.5400
	SD of ICSABI;ICSAB/5		(P)n/a	(PM)2.8530	(P)3.2580	(P)0.2275	(PI)313.7000	(M)649.5000	(P)0.0753	(P)0.0706	(P)0.3145	(P)0.9751
	%RSD of ICSABI;ICSAB/5		(P)n/a	(PM)0.1377	(P)17.3400	(P)0.5871	(PI)4.8030	(M)0.7212	(P)0.2727	(P)0.3436	(P)1.2430	(P)3.4170
1	CCV1;CCV	10/29/2005 12:01:12	(P)1.046	(P)98.1100	(P)99.3000	(P)99.8900	(PI)3.3660	(P)996.7000	(P)99.6800	(P)99.1000	(P)98.2000	(P)100.5000
2	CCV1;CCV	10/29/2005 12:02:08	(P)1.029	(P)99.1400	(P)98.0300	(P)98.3700	(PI)21.5700	(P)992.0000	(P)99.8100	(P)98.6900	(P)98.7000	(P)99.3900
3	CCV1;CCV	10/29/2005 12:03:04	(P)1.033	(P)99.6500	(P)98.3600	(P)98.6400	(PI)20.8800	(P)986.2000	(P)99.0300	(P)98.7900	(P)98.5800	(P)99.5700
	Mean of CCV1;CCV	10/29/2005 12:01:12	(P)1.046	(P)98.9700	(P)98.5700	(P)98.9700	(PI)15.2700	(P)991.6000	(P)99.5100	(P)98.8600	(P)98.4900	(P)99.8100
	SD of CCV1;CCV		(P)n/a	(P)0.7817	(P)0.6615	(P)0.8107	(PI)10.3100	(P)5.2300	(P)0.4180	(P)0.2121	(P)0.2576	(P)0.5824
	%RSD of CCV1;CCV		(P)n/a	(P)0.7899	(P)0.6712	(P)0.8192	(PI)67.5500	(P)0.5274	(P)0.4200	(P)0.2145	(P)0.2615	(P)0.5835
1	CCB1;CCB	10/29/2005 12:06:57	(P)1.011	(P)0.0012	(P)-0.0058	(P)0.0025	(PI)-57.9800	(P)0.2448	(P)-0.0053	(P)-0.0022	(P)-0.0086	(P)0.0180
2	CCB1;CCB	10/29/2005 12:07:53	(P)1.015	(P)0.0282	(P)-0.0520	(P)-0.0058	(PI)-52.5200	(P)-1.0340	(P)-0.0092	(P)-0.0057	(P)-0.0058	(P)0.0114
3	CCB1;CCB	10/29/2005 12:08:49	(P)1.022	(P)0.1200	(P)0.0703	(P)0.0136	(PI)-39.8400	(P)0.6612	(P)0.0189	(P)0.0068	(P)0.0039	(P)0.0265

Excluded In Calib	Excluded In Results	Peak Run	Excluded	Multi Element	Semi Quant	Internal Standard	Standard Addition					
Dilution Corrected Concentration Per Analyte			S-Calibration Has Edited Standard F-Interference Correction Failed		E-Calibration Edited T-Tripped	I-Invalid Calibration P-Pulse Counting	V-Valley Integration Failed M-Result Over Max					
Run	Label	TimeStamp	66Zn	75As	76Se	77Ar Cl	78Se	83Kr	89Y	95Mo	103Rh	106Pd
			ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb
	SD of ICB;ICB		(P)0.0530	(P)0.0916	(P)0.0000	(P)0.0000	(P)0.2835	(P)0.0000	(P)n/a	(P)0.0015	(P)n/a	(P)0.0000
	%RSD of ICB;ICB		(P)74.3300	(P)273.1000	(P)0.0000	(P)0.0000	(P)940.6000	(P)0.0000	(P)n/a	(P)11.5900	(P)n/a	(P)0.0000
1	CR11;CRI	10/28/2005 11:43:59	(P)2.1160	(P)1.3100	(P)0.0000	(P)0.0000	(P)5.0540	(P)0.0000	(P)1.019	(P)-0.0009	(P)1.010	(P)0.0000
2	CR11;CRI	10/28/2005 11:44:55	(P)1.9840	(P)0.9653	(P)0.0000	(P)0.0000	(P)5.3880	(P)0.0000	(P)1.024	(P)0.0059	(P)1.007	(P)0.0000
3	CR11;CRI	10/28/2005 11:45:51	(P)1.9970	(P)0.8758	(P)0.0000	(P)0.0000	(P)4.9740	(P)0.0000	(P)1.024	(P)0.0088	(P)1.010	(P)0.0000
	Mean of CR11;CRI	10/28/2005 11:43:59	(P)2.0320	(P)1.0500	(P)0.0000	(P)0.0000	(P)5.1390	(P)0.0000	(P)1.019	(P)0.0046	(P)1.010	(P)0.0000
	SD of CR11;CRI		(P)0.0726	(P)0.2295	(P)0.0000	(P)0.0000	(P)0.2196	(P)0.0000	(P)n/a	(P)0.0049	(P)n/a	(P)0.0000
	%RSD of CR11;CRI		(P)3.5720	(P)21.8500	(P)0.0000	(P)0.0000	(P)4.2730	(P)0.0000	(P)n/a	(P)107.8000	(P)n/a	(P)0.0000
1	ICSA1;ICSA/5	10/28/2005 11:49:43	(P)15.9900	(P)0.9251	(P)0.0000	(P)0.0000	(P)-0.4078	(P)0.0000	(P)1.108	(PM)1863.000	(P)1.134	(P)0.0000
2	ICSA1;ICSA/5	10/28/2005 11:50:39	(P)15.1300	(P)0.7127	(P)0.0000	(P)0.0000	(P)-0.3837	(P)0.0000	(P)1.110	(PM)1866.000	(P)1.134	(P)0.0000
3	ICSA1;ICSA/5	10/28/2005 11:51:35	(P)14.7700	(P)2.0660	(P)0.0000	(P)0.0000	(P)1.0530	(P)0.0000	(P)1.085	(PM)1847.000	(P)1.108	(P)0.0000
	Mean of ICSA1;ICSA/5	10/28/2005 11:49:43	(P)15.3000	(P)1.2350	(P)0.0000	(P)0.0000	(P)0.0870	(P)0.0000	(P)1.108	(PM)1858.000	(P)1.134	(P)0.0000
	SD of ICSA1;ICSA/5		(P)0.6318	(P)0.7277	(P)0.0000	(P)0.0000	(P)0.8363	(P)0.0000	(P)n/a	(PM)10.2400	(P)n/a	(P)0.0000
	%RSD of ICSA1;ICSA		(P)4.1300	(P)58.9500	(P)0.0000	(P)0.0000	(P)961.1000	(P)0.0000	(P)n/a	(PM)0.5511	(P)n/a	(P)0.0000
1	ICSAB1;ICSAB/5	10/28/2005 11:55:27	(P)32.5900	(P)19.9700	(P)0.0000	(P)0.0000	(P)19.1900	(P)0.0000	(P)1.120	(PM)1913.000	(P)1.137	(P)0.0000
2	ICSAB1;ICSAB/5	10/28/2005 11:56:23	(P)30.8500	(P)19.8100	(P)0.0000	(P)0.0000	(P)22.3300	(P)0.0000	(P)1.121	(PM)1931.000	(P)1.144	(P)0.0000
3	ICSAB1;ICSAB/5	10/28/2005 11:57:19	(P)32.1300	(P)20.2500	(P)0.0000	(P)0.0000	(P)18.3500	(P)0.0000	(P)1.107	(PM)1941.000	(P)1.119	(P)0.0000
	Mean of ICSAB1;ICSAB/5	10/28/2005 11:55:27	(P)31.8600	(P)20.0100	(P)0.0000	(P)0.0000	(P)19.9600	(P)0.0000	(P)1.120	(PM)1929.000	(P)1.137	(P)0.0000
	SD of ICSAB1;ICSAB/5		(P)0.8989	(P)0.2231	(P)0.0000	(P)0.0000	(P)2.0970	(P)0.0000	(P)n/a	(PM)14.0700	(P)n/a	(P)0.0000
	%RSD of ICSAB1;ICSAB/5		(P)2.8220	(P)1.1150	(P)0.0000	(P)0.0000	(P)10.5100	(P)0.0000	(P)n/a	(PM)0.7296	(P)n/a	(P)0.0000
1	CCV1;CCV	10/29/2005 12:01:12	(P)100.5000	(P)99.1600	(P)0.0000	(P)0.0000	(P)98.1000	(P)0.0000	(P)1.046	(P)99.9100	(P)1.046	(P)0.0000
2	CCV1;CCV	10/29/2005 12:02:08	(P)100.0000	(P)99.4200	(P)0.0000	(P)0.0000	(P)96.4400	(P)0.0000	(P)1.031	(P)97.8300	(P)1.030	(P)0.0000
3	CCV1;CCV	10/29/2005 12:03:04	(P)99.5600	(P)99.2000	(P)0.0000	(P)0.0000	(P)99.1500	(P)0.0000	(P)1.058	(P)98.7200	(P)1.038	(P)0.0000
	Mean of CCV1;CCV	10/29/2005 12:01:12	(P)100.0000	(P)99.2600	(P)0.0000	(P)0.0000	(P)97.9000	(P)0.0000	(P)1.046	(P)98.8200	(P)1.046	(P)0.0000
	SD of CCV1;CCV		(P)0.4719	(P)0.1407	(P)0.0000	(P)0.0000	(P)1.3670	(P)0.0000	(P)n/a	(P)1.0430	(P)n/a	(P)0.0000
	%RSD of CCV1;CCV		(P)0.4718	(P)0.1418	(P)0.0000	(P)0.0000	(P)1.3960	(P)0.0000	(P)n/a	(P)1.0560	(P)n/a	(P)0.0000
1	CCB1;CCB	10/29/2005 12:06:57	(P)0.0435	(P)-0.0685	(P)0.0000	(P)0.0000	(P)0.0903	(P)0.0000	(P)1.011	(P)0.0439	(P)1.001	(P)0.0000
2	CCB1;CCB	10/29/2005 12:07:53	(P)0.0636	(P)0.0209	(P)0.0000	(P)0.0000	(P)-0.2190	(P)0.0000	(P)1.015	(P)0.0251	(P)1.016	(P)0.0000
3	CCB1;CCB	10/29/2005 12:08:49	(P)0.1022	(P)0.1047	(P)0.0000	(P)0.0000	(P)0.5273	(P)0.0000	(P)1.024	(P)0.0666	(P)1.010	(P)0.0000

Excluded In Calib	Excluded In Results	Peak Run	Excluded	Multi Element	Semi Quant	Internal Standard	Standard Addition
Dilution Corrected Concentration Per Analyte			S-Calibration Has Edited Standard F-Interference Correction Failed	E-Calibration Edited T-Tripped	I-Invalid Calibration P-Pulse Counting	V-Valley Integration Failed M-Result Over Max	

Run	Label	TimeStamp	107Ag	108Pd	111Cd	113In	115In	121Sb	123Sb	137Ba	159Tb	165Ho
			ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb
	SD of ICB;ICB		(P)0.0037	(P)0.0000	(P)0.0056	(P)0.0000	(P)n/a	(P)0.1056	(P)0.1011	(P)0.0073	(P)n/a	(P)n/a
	%RSD of ICB;ICB		(P)20.6400	(P)0.0000	(P)19.9900	(P)0.0000	(P)n/a	(P)28.3000	(P)27.8300	(P)63.8000	(P)n/a	(P)n/a
1	CRII;CRI	10/28/2005 11:43:59	(P)0.9993	(P)0.0000	(P)1.0170	(P)0.0000	(P)1.011	(P)2.0070	(P)2.1150	(P)9.4000	(P)1.018	(P)1.013
2	CRII;CRI	10/28/2005 11:44:55	(P)0.9895	(P)0.0000	(P)0.9021	(P)0.0000	(P)1.023	(P)2.0990	(P)1.9860	(P)9.7370	(P)1.034	(P)1.034
3	CRII;CRI	10/28/2005 11:45:51	(P)0.9728	(P)0.0000	(P)0.9419	(P)0.0000	(P)1.025	(P)2.0440	(P)2.0700	(P)9.5260	(P)1.011	(P)1.024
	Mean of CRII;CRI	10/28/2005 11:43:59	(P)0.9872	(P)0.0000	(P)0.9538	(P)0.0000	(P)1.011	(P)2.0500	(P)2.0570	(P)9.5540	(P)1.018	(P)1.013
	SD of CRII;CRI		(P)0.0134	(P)0.0000	(P)0.0586	(P)0.0000	(P)n/a	(P)0.0465	(P)0.0656	(P)0.1700	(P)n/a	(P)n/a
	%RSD of CRII;CRI		(P)1.3580	(P)0.0000	(P)6.1460	(P)0.0000	(P)n/a	(P)2.2690	(P)3.1880	(P)1.7800	(P)n/a	(P)n/a
1	ICSAI;ICSA/5	10/28/2005 11:49:43	(P)0.0769	(P)0.0000	(P)0.8707	(P)0.0000	(P)1.137	(P)1.4050	(P)1.3130	(P)1.2760	(P)1.081	(P)1.085
2	ICSAI;ICSA/5	10/28/2005 11:50:39	(P)0.0673	(P)0.0000	(P)0.4941	(P)0.0000	(P)1.132	(P)1.3740	(P)1.4260	(P)1.4720	(P)1.061	(P)1.056
3	ICSAI;ICSA/5	10/28/2005 11:51:35	(P)0.0686	(P)0.0000	(P)0.4440	(P)0.0000	(P)1.102	(P)1.2710	(P)1.3150	(P)1.3200	(P)1.070	(P)1.075
	Mean of ICSAI;ICSA/5	10/28/2005 11:49:43	(P)0.0709	(P)0.0000	(P)0.6029	(P)0.0000	(P)1.137	(P)1.3500	(P)1.3510	(P)1.3560	(P)1.081	(P)1.085
	SD of ICSAI;ICSA/5		(P)0.0052	(P)0.0000	(P)0.2333	(P)0.0000	(P)n/a	(P)0.0703	(P)0.0648	(P)0.1030	(P)n/a	(P)n/a
	%RSD of ICSAI;ICSA		(P)7.3050	(P)0.0000	(P)38.6900	(P)0.0000	(P)n/a	(P)5.2100	(P)4.7930	(P)7.5980	(P)n/a	(P)n/a
1	ICSABI;ICSAB/5	10/28/2005 11:55:27	(P)19.3900	(P)0.0000	(P)19.5200	(P)0.0000	(P)1.128	(P)21.4900	(P)21.5700	(P)21.1600	(P)1.082	(P)1.097
2	ICSABI;ICSAB/5	10/28/2005 11:56:23	(P)19.1400	(P)0.0000	(P)20.1000	(P)0.0000	(P)1.136	(P)21.8900	(P)21.0000	(P)20.9000	(P)1.091	(P)1.089
3	ICSABI;ICSAB/5	10/28/2005 11:57:19	(P)19.3900	(P)0.0000	(P)19.9900	(P)0.0000	(P)1.122	(P)22.0900	(P)22.0300	(P)21.4600	(P)1.072	(P)1.079
	Mean of ICSABI;ICSAB/5	10/28/2005 11:55:27	(P)19.3000	(P)0.0000	(P)19.8700	(P)0.0000	(P)1.128	(P)21.8200	(P)21.5300	(P)21.1700	(P)1.082	(P)1.097
	SD of ICSABI;ICSAB/5		(P)0.1433	(P)0.0000	(P)0.3084	(P)0.0000	(P)n/a	(P)0.3035	(P)0.5168	(P)0.2813	(P)n/a	(P)n/a
	%RSD of ICSABI;ICSAB/5		(P)0.7424	(P)0.0000	(P)1.5520	(P)0.0000	(P)n/a	(P)1.3910	(P)2.4000	(P)1.3280	(P)n/a	(P)n/a
1	CCV1;CCV	10/29/2005 12:01:12	(P)100.2000	(P)0.0000	(P)99.6600	(P)0.0000	(P)1.047	(P)98.3300	(P)98.5500	(P)98.7900	(P)1.015	(P)1.029
2	CCV1;CCV	10/29/2005 12:02:08	(P)100.0000	(P)0.0000	(P)99.2400	(P)0.0000	(P)1.042	(P)99.0700	(P)99.5800	(P)97.5600	(P)1.013	(P)1.032
3	CCV1;CCV	10/29/2005 12:03:04	(P)100.0000	(P)0.0000	(P)99.9800	(P)0.0000	(P)1.050	(P)99.7400	(P)99.8900	(P)100.3000	(P)1.032	(P)1.051
	Mean of CCV1;CCV	10/29/2005 12:01:12	(P)100.1000	(P)0.0000	(P)99.6300	(P)0.0000	(P)1.047	(P)99.0500	(P)99.3400	(P)98.8800	(P)1.015	(P)1.029
	SD of CCV1;CCV		(P)0.0894	(P)0.0000	(P)0.3702	(P)0.0000	(P)n/a	(P)0.7072	(P)0.7035	(P)1.3610	(P)n/a	(P)n/a
	%RSD of CCV1;CCV		(P)0.0894	(P)0.0000	(P)0.3716	(P)0.0000	(P)n/a	(P)0.7140	(P)0.7082	(P)1.3760	(P)n/a	(P)n/a
1	CCB1;CCB	10/29/2005 12:06:57	(P)0.0089	(P)0.0000	(P)-0.0237	(P)0.0000	(P)1.013	(P)0.2113	(P)0.2155	(P)0.0101	(P)0.999	(P)1.002
2	CCB1;CCB	10/29/2005 12:07:53	(P)0.0070	(P)0.0000	(P)-0.0006	(P)0.0000	(P)1.013	(P)0.1411	(P)0.1540	(P)0.0017	(P)1.003	(P)1.017
3	CCB1;CCB	10/29/2005 12:08:49	(P)0.0144	(P)0.0000	(P)-0.0036	(P)0.0000	(P)1.024	(P)0.1232	(P)0.1305	(P)0.0190	(P)1.015	(P)1.011

Excluded In Calib	Excluded In Results	Peak Run	Excluded	Multi Element	Semi Quant	Internal Standard	Standard Addition	
Dilution Corrected Concentration Per Analyte			S-Calibration Has Edited Standard F-Interference Correction Failed	E-Calibration Edited T-Tripped	I-Invalid Calibration P-Pulse Counting	V-Valley Integration Failed M-Result Over Max		
Run	Label	TimeStamp	203Tl	205Tl	206Pb	207Pb	208Pb	209Bi
			ppb	ppb	ppb	ppb	ppb	ppb
	SD of ICB;ICB		(P)0.0056	(P)0.0102	(P)0.0056	(P)0.0060	(P)0.0046	(P)n/a
	%RSD of ICB;ICB		(P)20.2100	(P)36.9000	(P)44.4000	(P)40.2600	(P)32.6500	(P)n/a
1	CRII;CRI	10/28/2005 11:43:59	(P)1.0160	(P)1.0180	(P)0.9525	(P)0.9650	(P)0.9666	(P)1.010
2	CRII;CRI	10/28/2005 11:44:55	(P)1.0100	(P)1.0370	(P)0.9614	(P)0.9691	(P)0.9713	(P)1.026
3	CRII;CRI	10/28/2005 11:45:51	(P)1.0390	(P)1.0360	(P)0.9984	(P)0.9811	(P)0.9685	(P)1.025
	Mean of CRII;CRI	10/28/2005 11:43:59	(P)1.0220	(P)1.0300	(P)0.9708	(P)0.9717	(P)0.9688	(P)1.010
	SD of CRII;CRI		(P)0.0150	(P)0.0103	(P)0.0243	(P)0.0083	(P)0.0024	(P)n/a
	%RSD of CRII;CRI		(P)1.4660	(P)1.0040	(P)2.5060	(P)0.8582	(P)0.2439	(P)n/a
1	ICSAI;ICSA/5	10/28/2005 11:49:43	(P)0.0326	(P)0.0449	(P)4.5650	(P)4.3210	(P)4.3190	(P)1.115
2	ICSAI;ICSA/5	10/28/2005 11:50:39	(P)0.0477	(P)0.0582	(P)4.6620	(P)4.4440	(P)4.4000	(P)1.102
3	ICSAI;ICSA/5	10/28/2005 11:51:35	(P)0.0667	(P)0.0664	(P)4.4920	(P)4.3610	(P)4.3320	(P)1.090
	Mean of ICSAI;ICSA/5	10/28/2005 11:49:43	(P)0.0490	(P)0.0565	(P)4.5730	(P)4.3750	(P)4.3500	(P)1.115
	SD of ICSAI;ICSA/5		(P)0.0171	(P)0.0108	(P)0.0856	(P)0.0625	(P)0.0439	(P)n/a
	%RSD of ICSAI;ICSA		(P)34.8800	(P)19.1600	(P)1.8720	(P)1.4290	(P)1.0080	(P)n/a
1	ICSABI;ICSAB/5	10/28/2005 11:55:27	(P)20.5500	(P)20.0600	(P)23.8900	(P)23.4700	(P)23.6000	(P)1.108
2	ICSABI;ICSAB/5	10/28/2005 11:56:23	(P)20.5200	(P)20.1800	(P)24.2500	(P)23.9800	(P)24.0600	(P)1.121
3	ICSABI;ICSAB/5	10/28/2005 11:57:19	(P)20.3300	(P)20.5300	(P)24.4400	(P)24.3300	(P)24.2700	(P)1.105
	Mean of ICSABI;ICSA	10/28/2005 11:55:27	(P)20.4700	(P)20.2600	(P)24.1900	(P)23.9300	(P)23.9800	(P)1.108
	SD of ICSABI;ICSAB/		(P)0.1196	(P)0.2468	(P)0.2767	(P)0.4353	(P)0.3422	(P)n/a
	%RSD of ICSABI;ICS		(P)0.5843	(P)1.2180	(P)1.1440	(P)1.8190	(P)1.4270	(P)n/a
1	CCV1;CCV	10/29/2005 12:01:12	(P)98.8000	(P)100.2000	(P)99.5300	(P)99.3600	(P)98.6500	(P)1.033
2	CCV1;CCV	10/29/2005 12:02:08	(P)98.4900	(P)99.1400	(P)98.8700	(P)98.5400	(P)98.7400	(P)1.024
3	CCV1;CCV	10/29/2005 12:03:04	(P)98.9300	(P)100.1000	(P)99.2700	(P)99.1200	(P)99.1300	(P)1.036
	Mean of CCV1;CCV	10/29/2005 12:01:12	(P)98.7400	(P)99.8100	(P)99.2300	(P)99.0100	(P)98.8400	(P)1.033
	SD of CCV1;CCV		(P)0.2297	(P)0.5878	(P)0.3325	(P)0.4197	(P)0.2589	(P)n/a
	%RSD of CCV1;CCV		(P)0.2326	(P)0.5889	(P)0.3351	(P)0.4239	(P)0.2619	(P)n/a
1	CCB1;CCB	10/29/2005 12:06:57	(P)0.0111	(P)0.0104	(P)-0.0012	(P)-0.0011	(P)-0.0006	(P)1.002
2	CCB1;CCB	10/29/2005 12:07:53	(P)0.0049	(P)0.0073	(P)-0.0036	(P)-0.0042	(P)-0.0037	(P)0.996
3	CCB1;CCB	10/29/2005 12:08:49	(P)0.0162	(P)0.0173	(P)0.0111	(P)0.0151	(P)0.0118	(P)1.010

Excluded In Calib		Excluded In Results	Peak Run	Excluded	Multi Element	Semi Quant	Internal Standard	Standard Addition				
Dilution Corrected Concentration Per Analyte			S-Calibration Has Edited Standard F-Interference Correction Failed		E-Calibration Edited T-Tripped	I-Invalid Calibration P-Pulse Counting	V-Valley Integration Failed M-Result Over Max					
Run	Label	TimeStamp	6Li	7Li	9Be	23Na	24Mg	25Mg	26Mg	27Al	39K	44Ca
			ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb
	Mean of CCB1;CCB	10/29/2005 12:06:57	(P)0.984	(P)0.0000	(P)-0.0225	(P)-2.6060	(P)0.9171	(P)0.9499	(P)0.8526	(P)0.9868	(P)1.8790	(P)1.5840
	SD of CCB1;CCB		(P)n/a	(P)0.0000	(P)0.0126	(P)1.7390	(P)0.7147	(P)0.6896	(P)0.7100	(P)0.5623	(P)0.9756	(P)0.9747
	%RSD of CCB1;CCB		(P)n/a	(P)0.0000	(P)56.0800	(P)66.7300	(P)77.9300	(P)72.6000	(P)83.2700	(P)56.9800	(P)51.9300	(P)61.5400
1	PBW;PBW	10/29/2005 12:12:41	(P)0.961	(P)0.0000	(P)-0.0354	(P)267.3000	(P)4.5660	(P)4.5190	(P)4.9660	(P)70.6400	(P)80.2400	(P)18.0100
2	PBW;PBW	10/29/2005 12:13:37	(P)0.972	(P)0.0000	(P)-0.0322	(P)283.8000	(P)4.8540	(P)4.5210	(P)4.6890	(P)74.7100	(P)75.7200	(P)18.6000
3	PBW;PBW	10/29/2005 12:14:33	(P)0.978	(P)0.0000	(P)-0.0189	(P)306.9000	(P)5.1820	(P)4.7250	(P)5.0160	(P)79.0600	(P)79.1600	(P)20.4800
	Mean of PBW;PBW	10/29/2005 12:12:41	(P)0.961	(P)0.0000	(P)-0.0288	(P)286.0000	(P)4.8670	(P)4.5880	(P)4.8910	(P)74.8000	(P)78.3700	(P)19.0300
	SD of PBW;PBW		(P)n/a	(P)0.0000	(P)0.0087	(P)19.8900	(P)0.3086	(P)0.1183	(P)0.1759	(P)4.2110	(P)2.3580	(P)1.2880
	%RSD of PBW;PBW		(P)n/a	(P)0.0000	(P)30.3500	(P)6.9560	(P)6.3400	(P)2.5790	(P)3.5970	(P)5.6300	(P)3.0090	(P)6.7690
1	LCSW;LCSW	10/29/2005 12:18:19	(P)1.020	(P)0.0000	(P)81.1700	(M)2290.0000	(P)988.8000	(P)984.6000	(P)993.8000	(PM)553.7000	1830.0000	(P)1759.0000
2	LCSW;LCSW	10/29/2005 12:19:15	(P)1.019	(P)0.0000	(P)81.4200	(M)2259.0000	(P)1001.0000	(P)988.3000	(P)989.8000	(PM)557.1000	1825.0000	(P)1758.0000
3	LCSW;LCSW	10/29/2005 12:20:11	(P)1.006	(P)0.0000	(P)79.9200	(M)2248.0000	(P)984.7000	(P)988.6000	(P)978.3000	(PM)558.8000	1809.0000	(P)1760.0000
	Mean of LCSW;LCSW	10/29/2005 12:18:19	(P)1.020	(P)0.0000	(P)80.8400	(M)2265.0000	(P)991.5000	(P)987.1000	(P)987.3000	(PM)556.5000	1821.0000	(P)1759.0000
	SD of LCSW;LCSW		(P)n/a	(P)0.0000	(P)0.8057	(M)22.0400	(P)8.4690	(P)2.2300	(P)8.0560	(PM)2.6120	10.8200	(P)0.5534
	%RSD of LCSW;LCS		(P)n/a	(P)0.0000	(P)0.9967	(M)0.9729	(P)0.8542	(P)0.2259	(P)0.8159	(PM)0.4694	0.5939	(P)0.0315
1	MH1FZ4L;ISD/5	10/29/2005 12:23:57	(P)0.987	(P)0.0000	(P)0.0102	1197.0000	(P)758.3000	(P)769.8000	(P)760.7000	(P)119.2000	(P)304.9000	(PM)4582.000
2	MH1FZ4L;ISD/5	10/29/2005 12:24:53	(P)1.002	(P)0.0000	(P)-0.0175	1205.0000	(P)769.4000	(P)766.5000	(P)765.6000	(P)120.6000	(P)307.7000	(PM)4620.000
3	MH1FZ4L;ISD/5	10/29/2005 12:25:49	(P)0.984	(P)0.0000	(P)-0.0317	1217.0000	(P)761.0000	(P)760.8000	(P)762.2000	(P)120.0000	(P)313.4000	(PM)4617.000
	Mean of MH1FZ4L;ISC	10/29/2005 12:23:57	(P)0.987	(P)0.0000	(P)-0.0130	1206.0000	(P)762.9000	(P)765.7000	(P)762.8000	(P)119.9000	(P)308.6000	(PM)4606.000
	SD of MH1FZ4L;ISD/5		(P)n/a	(P)0.0000	(P)0.0213	10.0200	(P)5.7980	(P)4.5780	(P)2.5350	(P)0.7132	(P)4.3340	(PM)21.0100
	%RSD of MH1FZ4L;I		(P)n/a	(P)0.0000	(P)164.2000	0.8311	(P)0.7600	(P)0.5979	(P)0.3323	(P)0.5947	(P)1.4040	(PM)0.4562
1	MH1FZ4;S	10/29/2005 12:29:36	(P)0.973	(P)0.0000	(P)-0.0272	(M)4375.0000	(M)4002.0000	(PM)4018.000	(PM)4012.000	(P)179.3000	1218.0000	(M)25130.000
2	MH1FZ4;S	10/29/2005 12:30:32	(P)0.981	(P)0.0000	(P)-0.0088	(M)4374.0000	(M)4050.0000	(PM)4043.000	(PM)4047.000	(P)173.4000	1219.0000	(M)25190.000
3	MH1FZ4;S	10/29/2005 12:31:28	(P)0.980	(P)0.0000	(P)-0.0191	(M)4437.0000	(M)4109.0000	(PM)4073.000	(PM)4057.000	(P)173.2000	1229.0000	(M)25690.000
	Mean of MH1FZ4;S	10/29/2005 12:29:36	(P)0.973	(P)0.0000	(P)-0.0184	(M)4395.0000	(M)4054.0000	(PM)4044.000	(PM)4038.000	(P)175.3000	1222.0000	(M)25340.000
	SD of MH1FZ4;S		(P)n/a	(P)0.0000	(P)0.0092	(M)36.1000	(M)53.6200	(PM)27.7300	(PM)23.5200	(P)3.4730	6.3190	(M)308.3000
	%RSD of MH1FZ4;S		(P)n/a	(P)0.0000	(P)50.3600	(M)0.8214	(M)1.3230	(PM)0.6857	(PM)0.5823	(P)1.9810	0.5171	(M)1.2170
1	MH1FZ4D;S2	10/29/2005 12:35:16	(P)0.966	(P)0.0000	(P)-0.0123	(M)4645.0000	(M)4285.0000	(PM)4222.000	(PM)4268.000	(P)177.5000	1298.0000	(M)26840.000
2	MH1FZ4D;S2	10/29/2005 12:36:12	(P)0.969	(P)0.0000	(P)-0.0172	(M)4631.0000	(M)4333.0000	(PM)4252.000	(PM)4225.000	(P)178.2000	1292.0000	(M)26830.000

Excluded In Calib		Excluded In Results	Peak Run	Excluded	Multi Element	Semi Quant	Internal Standard	Standard Addition				
Dilution Corrected Concentration Per Analyte			S-Calibration Has Edited Standard		E-Calibration Edited		I-Invalid Calibration	V-Valley Integration Failed				
			F-Interference Correction Failed		T-Tripped		P-Pulse Counting	M-Result Over Max				
Run	Label	TimeStamp	45Sc	47Ti	51V	52Cr	53Cr	54Fe	55Mn	59Co	60Ni	65Cu
			ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb
	Mean of CCB1;CCB	10/29/2005 12:06:57	(P)1.011	(P)0.0498	(P)0.0042	(P)0.0034	(PI)-50.1100	(P)-0.0425	(P)0.0015	(P)-0.0004	(P)-0.0035	(P)0.0186
	SD of CCB1;CCB		(P)n/a	(P)0.0622	(P)0.0617	(P)0.0097	(PI)9.3070	(P)0.8833	(P)0.0152	(P)0.0065	(P)0.0065	(P)0.0076
	%RSD of CCB1;CCB		(P)n/a	(P)124.9000	(P)1474.0000	(P)282.4000	(PI)18.5700	(P)2076.0000	(P)1019.0000	(P)1843.0000	(P)186.7000	(P)40.7100
1	PBW;PBW	10/29/2005 12:12:41	(P)1.015	(P)0.2616	(P)0.0569	(P)0.0779	(PI)-36.2100	(P)0.4988	(P)0.0355	(P)0.0019	(P)0.0414	(P)0.3357
2	PBW;PBW	10/29/2005 12:13:37	(P)1.013	(P)0.2291	(P)-0.0377	(P)0.0673	(PI)-30.3400	(P)1.6600	(P)0.0162	(P)0.0043	(P)0.0231	(P)0.3408
3	PBW;PBW	10/29/2005 12:14:33	(P)1.028	(P)0.3146	(P)0.0108	(P)0.0751	(PI)-35.7700	(P)0.6799	(P)0.0489	(P)0.0042	(P)0.0186	(P)0.4537
	Mean of PBW;PBW	10/29/2005 12:12:41	(P)1.015	(P)0.2684	(P)0.0100	(P)0.0735	(PI)-34.1000	(P)0.9464	(P)0.0335	(P)0.0035	(P)0.0277	(P)0.3767
	SD of PBW;PBW		(P)n/a	(P)0.0431	(P)0.0473	(P)0.0055	(PI)3.2680	(P)0.6249	(P)0.0165	(P)0.0013	(P)0.0121	(P)0.0667
	%RSD of PBW;PBW		(P)n/a	(P)16.0700	(P)472.1000	(P)7.4400	(PI)9.5830	(P)66.0400	(P)49.1700	(P)38.3600	(P)43.6800	(P)17.7100
1	LCSW;LCSW	10/29/2005 12:18:19	(P)1.043	(P)0.5448	(P)82.4400	(P)80.0200	(PI)102.6000	(P)838.0000	(P)81.7800	(P)81.3400	(P)80.2700	(P)83.1100
2	LCSW;LCSW	10/29/2005 12:19:15	(P)1.044	(P)0.6109	(P)82.1700	(P)80.7100	(PI)165.5000	(P)836.7000	(P)82.0800	(P)81.9800	(P)82.0100	(P)82.7900
3	LCSW;LCSW	10/29/2005 12:20:11	(P)1.046	(P)0.6284	(P)82.0600	(P)80.5500	(PI)66.9900	(P)828.2000	(P)81.7500	(P)81.8000	(P)80.3100	(P)82.4400
	Mean of LCSW;LCSW	10/29/2005 12:18:19	(P)1.043	(P)0.5947	(P)82.2200	(P)80.4200	(PI)111.7000	(P)834.3000	(P)81.8700	(P)81.7000	(P)80.8600	(P)82.7800
	SD of LCSW;LCSW		(P)n/a	(P)0.0440	(P)0.1926	(P)0.3581	(PI)49.9000	(P)5.2950	(P)0.1851	(P)0.3297	(P)0.9961	(P)0.3323
	%RSD of LCSW;LCS		(P)n/a	(P)7.4070	(P)0.2342	(P)0.4453	(PI)44.6700	(P)0.6347	(P)0.2262	(P)0.4035	(P)1.2320	(P)0.4014
1	MH1FZ4L;ISD/5	10/29/2005 12:23:57	(P)1.004	(P)0.3100	(P)0.0697	(P)0.1344	(PI)-19.8000	(P)3.1840	(P)0.9067	(P)0.2035	(P)0.1927	(P)0.2489
2	MH1FZ4L;ISD/5	10/29/2005 12:24:53	(P)1.001	(P)0.4454	(P)0.0396	(P)0.1610	(PI)-20.3900	(P)3.2990	(P)0.9072	(P)0.2003	(P)0.1657	(P)0.2239
3	MH1FZ4L;ISD/5	10/29/2005 12:25:49	(P)0.997	(P)0.3745	(P)0.0049	(P)0.1326	(PI)-26.9000	(P)4.0180	(P)0.9110	(P)0.1836	(P)0.1690	(P)0.2371
	Mean of MH1FZ4L;ISC	10/29/2005 12:23:57	(P)1.004	(P)0.3766	(P)0.0381	(P)0.1427	(PI)-22.3600	(P)3.5000	(P)0.9083	(P)0.1958	(P)0.1758	(P)0.2366
	SD of MH1FZ4L;ISD/5		(P)n/a	(P)0.0677	(P)0.0324	(P)0.0159	(PI)3.9380	(P)0.4524	(P)0.0024	(P)0.0107	(P)0.0147	(P)0.0125
	%RSD of MH1FZ4L;I		(P)n/a	(P)17.9900	(P)85.1400	(P)11.1600	(PI)17.6100	(P)12.9300	(P)0.2598	(P)5.4760	(P)8.3790	(P)5.2930
1	MH1FZ4;S	10/29/2005 12:29:36	(P)0.975	(P)1.5600	(P)0.2853	(P)0.2610	(PI)-29.2700	(P)50.7300	(P)4.7740	(P)1.0550	(P)0.8874	(P)0.9671
2	MH1FZ4;S	10/29/2005 12:30:32	(P)0.955	(P)1.2940	(P)0.2129	(P)0.2639	(PI)-42.8100	(P)46.0300	(P)4.6320	(P)1.0190	(P)0.7836	(P)0.9374
3	MH1FZ4;S	10/29/2005 12:31:28	(P)0.965	(P)1.2980	(P)0.2668	(P)0.2784	(PI)-29.6700	(P)47.7800	(P)4.7400	(P)1.0370	(P)0.9304	(P)0.9660
	Mean of MH1FZ4;S	10/29/2005 12:29:36	(P)0.975	(P)1.3840	(P)0.2550	(P)0.2678	(PI)-33.9200	(P)48.1800	(P)4.7150	(P)1.0370	(P)0.8671	(P)0.9568
	SD of MH1FZ4;S		(P)n/a	(P)0.1519	(P)0.0376	(P)0.0093	(PI)7.7030	(P)2.3740	(P)0.0740	(P)0.0179	(P)0.0755	(P)0.0168
	%RSD of MH1FZ4;S		(P)n/a	(P)10.9800	(P)14.7500	(P)3.4770	(PI)22.7100	(P)4.9260	(P)1.5690	(P)1.7220	(P)8.7020	(P)1.7580
1	MH1FZ4D;S2	10/29/2005 12:35:16	(P)0.970	(P)0.9605	(P)0.2400	(P)0.2747	(PI)-33.2600	(P)49.8500	(P)4.9780	(P)1.0530	(P)0.8970	(P)1.8200
2	MH1FZ4D;S2	10/29/2005 12:36:12	(P)0.972	(P)1.4280	(P)0.2567	(P)0.2836	(PI)-36.1500	(P)49.1500	(P)4.9940	(P)1.0700	(P)0.9669	(P)1.7680

Excluded In Calib		Excluded In Results	Peak Run	Excluded	Multi Element	Semi Quant	Internal Standard	Standard Addition				
Dilution Corrected Concentration Per Analyte			S-Calibration Has Edited Standard F-Interference Correction Failed		E-Calibration Edited T-Tripped	I-Invalid Calibration P-Pulse Counting	V-Valley Integration Failed M-Result Over Max					
Run	Label	TimeStamp	66Zn	75As	76Se	77Ar Cl	78Se	83Kr	89Y	95Mo	103Rh	106Pd
			ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb	ppb
	Mean of CCB1;CCB	10/29/2005 12:06:57	(P)0.0698	(P)0.0190	(P)0.0000	(P)0.0000	(P)0.1329	(P)0.0000	(P)1.011	(P)0.0452	(P)1.001	(P)0.0000
	SD of CCB1;CCB		(P)0.0298	(P)0.0866	(P)0.0000	(P)0.0000	(P)0.3750	(P)0.0000	(P)n/a	(P)0.0208	(P)n/a	(P)0.0000
	%RSD of CCB1;CCB		(P)42.7300	(P)455.0000	(P)0.0000	(P)0.0000	(P)282.2000	(P)0.0000	(P)n/a	(P)46.0100	(P)n/a	(P)0.0000
1	PBW;PBW	10/29/2005 12:12:41	(P)0.6371	(P)0.1439	(P)0.0000	(P)0.0000	(P)0.0178	(P)0.0000	(P)1.009	(P)0.0343	(P)1.002	(P)0.0000
2	PBW;PBW	10/29/2005 12:13:37	(P)0.8031	(P)0.0093	(P)0.0000	(P)0.0000	(P)0.6674	(P)0.0000	(P)1.015	(P)0.0422	(P)1.006	(P)0.0000
3	PBW;PBW	10/29/2005 12:14:33	(P)0.7663	(P)-0.0454	(P)0.0000	(P)0.0000	(P)-0.1323	(P)0.0000	(P)1.018	(P)0.0356	(P)1.010	(P)0.0000
	Mean of PBW;PBW	10/29/2005 12:12:41	(P)0.7355	(P)0.0359	(P)0.0000	(P)0.0000	(P)0.1843	(P)0.0000	(P)1.009	(P)0.0374	(P)1.002	(P)0.0000
	SD of PBW;PBW		(P)0.0872	(P)0.0974	(P)0.0000	(P)0.0000	(P)0.4251	(P)0.0000	(P)n/a	(P)0.0042	(P)n/a	(P)0.0000
	%RSD of PBW;PBW		(P)11.8600	(P)271.1000	(P)0.0000	(P)0.0000	(P)230.6000	(P)0.0000	(P)n/a	(P)11.2700	(P)n/a	(P)0.0000
1	LCSW;LCSW	10/29/2005 12:18:19	(P)167.8000	(P)160.1000	(P)0.0000	(P)0.0000	(P)161.8000	(P)0.0000	(P)1.048	(P)0.0360	(P)1.047	(P)0.0000
2	LCSW;LCSW	10/29/2005 12:19:15	(P)162.8000	(P)159.9000	(P)0.0000	(P)0.0000	(P)163.4000	(P)0.0000	(P)1.052	(P)0.0312	(P)1.055	(P)0.0000
3	LCSW;LCSW	10/29/2005 12:20:11	(P)164.9000	(P)157.2000	(P)0.0000	(P)0.0000	(P)161.7000	(P)0.0000	(P)1.044	(P)0.0429	(P)1.051	(P)0.0000
	Mean of LCSW;LCSW	10/29/2005 12:18:19	(P)165.2000	(P)159.1000	(P)0.0000	(P)0.0000	(P)162.3000	(P)0.0000	(P)1.048	(P)0.0367	(P)1.047	(P)0.0000
	SD of LCSW;LCSW		(P)2.5180	(P)1.6110	(P)0.0000	(P)0.0000	(P)0.9428	(P)0.0000	(P)n/a	(P)0.0059	(P)n/a	(P)0.0000
	%RSD of LCSW;LCS		(P)1.5240	(P)1.0130	(P)0.0000	(P)0.0000	(P)0.5810	(P)0.0000	(P)n/a	(P)15.9900	(P)n/a	(P)0.0000
1	MH1FZ4L;ISD/5	10/29/2005 12:23:57	(P)1.1000	(P)0.1812	(P)0.0000	(P)0.0000	(P)-0.0872	(P)0.0000	(P)1.004	(P)0.1165	(P)1.016	(P)0.0000
2	MH1FZ4L;ISD/5	10/29/2005 12:24:53	(P)1.1520	(P)0.0182	(P)0.0000	(P)0.0000	(P)0.0962	(P)0.0000	(P)1.013	(P)0.1215	(P)1.007	(P)0.0000
3	MH1FZ4L;ISD/5	10/29/2005 12:25:49	(P)0.9488	(P)0.0279	(P)0.0000	(P)0.0000	(P)-0.2910	(P)0.0000	(P)0.980	(P)0.1256	(P)1.006	(P)0.0000
	Mean of MH1FZ4L;ISC	10/29/2005 12:23:57	(P)1.0670	(P)0.0757	(P)0.0000	(P)0.0000	(P)-0.0940	(P)0.0000	(P)1.004	(P)0.1212	(P)1.016	(P)0.0000
	SD of MH1FZ4L;ISD/5		(P)0.1057	(P)0.0914	(P)0.0000	(P)0.0000	(P)0.1937	(P)0.0000	(P)n/a	(P)0.0045	(P)n/a	(P)0.0000
	%RSD of MH1FZ4L;I		(P)9.9060	(P)120.7000	(P)0.0000	(P)0.0000	(P)206.0000	(P)0.0000	(P)n/a	(P)3.7350	(P)n/a	(P)0.0000
1	MH1FZ4;S	10/29/2005 12:29:36	(P)3.9510	(P)0.4028	(P)0.0000	(P)0.0000	(P)0.0692	(P)0.0000	(P)1.004	(P)0.5305	(P)1.011	(P)0.0000
2	MH1FZ4;S	10/29/2005 12:30:32	(P)3.6700	(P)0.4444	(P)0.0000	(P)0.0000	(P)0.1185	(P)0.0000	(P)0.972	(P)0.5268	(P)0.996	(P)0.0000
3	MH1FZ4;S	10/29/2005 12:31:28	(P)3.8220	(P)0.5058	(P)0.0000	(P)0.0000	(P)0.2679	(P)0.0000	(P)0.978	(P)0.5588	(P)1.008	(P)0.0000
	Mean of MH1FZ4;S	10/29/2005 12:29:36	(P)3.8140	(P)0.4510	(P)0.0000	(P)0.0000	(P)0.1518	(P)0.0000	(P)1.004	(P)0.5387	(P)1.011	(P)0.0000
	SD of MH1FZ4;S		(P)0.1407	(P)0.0518	(P)0.0000	(P)0.0000	(P)0.1035	(P)0.0000	(P)n/a	(P)0.0175	(P)n/a	(P)0.0000
	%RSD of MH1FZ4;S		(P)3.6890	(P)11.4900	(P)0.0000	(P)0.0000	(P)68.1500	(P)0.0000	(P)n/a	(P)3.2420	(P)n/a	(P)0.0000
1	MH1FZ4D;S2	10/29/2005 12:35:16	(P)4.5420	(P)0.3814	(P)0.0000	(P)0.0000	(P)0.3698	(P)0.0000	(P)0.982	(P)0.5836	(P)1.005	(P)0.0000
2	MH1FZ4D;S2	10/29/2005 12:36:12	(P)4.8280	(P)0.4403	(P)0.0000	(P)0.0000	(P)-0.2956	(P)0.0000	(P)1.007	(P)0.5867	(P)1.018	(P)0.0000